



DEPARTMENT OF HEALTH & HUMAN SERVICES

Centers for Disease Control
and Prevention (CDC)

NIOSH Reference: TN-24896
Mfr. Reference: BNX004

National Institute for Occupational
Safety and Health (NIOSH)
National Personal Protective
Technology Laboratory (NPPTL)
626 Cochran's Mill Road
Pittsburgh, PA 15236-0070
Phone: 412-386-4000
Fax: 412-386-4051
July 19, 2021

Simon Ma
Chief Technology Officer
BNX Converting, LLC
16727 Park Row
Houston, TX 77084

Dear Simon Ma:

The National Institute for Occupational Safety and Health (NIOSH) has reviewed your request accepted May 10, 2021. This request was for approval of the F95W and F95B N95 air-purifying filtering facepiece respirators. The respirators, identified by the color of the outer fabric, are part numbers F95W (white), and F95B (black). The only difference is the color. The complete respirator configurations are detailed on assembly matrix file name *1114AMa.xls*, revision A, dated: 4/21/2021.

This request is granted. Approvals are granted only for documentation written in the English language. It is the manufacturer's responsibility to correctly translate materials desired to be produced in languages other than English. Approval number TC-84A-9362 has been assigned. This respirator is approved for protection against particulates at N95 filter efficiency level.

The approval label is included as an attachment to this letter. The abbreviated labels have been accepted as submitted. The cautions and limitations are listed on the approval label. Only those assemblies affected by this request, or where new approval numbers are assigned, apply to this approval action. Production approval labels cannot include information on unapproved configurations.

The approved assembly consists of the parts as listed on the approval label and the assembly matrix. Parts are to be marked with the numbers indicated on the approval label in a legible and permanent manner (marking cannot be removed without evidence of its previous presence).

This certificate of approval is not an endorsement of the respirator by NIOSH, and such endorsement shall not be stated or implied in advertisements or other publicity. However, you may publicize the fact that this respirator has met the requirements of Title 42, *Code of Federal Regulations*, Part 84 (42 CFR 84).

Page 2- Simon Ma- TN-24896

No changes may be made to any respirators and accompanying documentation without prior written approval of NIOSH. Requests for changes must be submitted to NIOSH and a modification of this approval must be granted before changes are made.

Sincerely,

Jeffrey Peterson
Chief, Conformity Verification and
Standards Development Branch

Enclosures